


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;">Accredited to ISO/IEC 17025:2005</p>	<h3 style="margin: 0;">VEGA Environmental Consultants Ltd</h3> <p style="margin: 0;">Issue No: 013 Issue date: 03 September 2007</p>	
	<p>Church Buildings Station Road Seaton Delaval Whitley Bay Tyne and Wear NE25 0PT</p>	<p>Contact: Mr E Brown Tel: +44 (0)191-298 0198 Fax: +44 (0)191-298 0199 E-Mail: edbrown@vegaec.co.uk Website: www.vegaec.co.uk</p>

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Church Buildings Station Road Seaton Delaval Whitley Bay Tyne and Wear NE25 0PT</p> <p style="text-align: right;">Local contact Mr E Brown</p>	Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B
Mobile laboratories		C



1634
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

VEGA Environmental Consultants Ltd
Issue No: 013 Issue date: 03 September 2007

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248 Guidance: February 2005 and Documented In-House Procedure ASB/1 using Membrane Filter Method and Phase Contrast Microscopy (PCM)	A, B, C
	Sampling of air for fibre counting	HSG 248 Guidance: February 2005 and Documented In-House Procedure ASB/10	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 and Documented In-House Procedure ASB/24 Membrane Filter Method using Phase Contrast Microscopy (PCM)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248 Guidance: February 2005 and Documented In-House Procedure No ASB/3, using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for asbestos identification	HSG 248 Guidance: February 2005 and Documented In-House Method No ASB/6	B
ENVIRONMENTAL SAMPLES	<u>Health and Hygiene</u>		
Cleanrooms and Associated Environments	Sampling of air for particulate counting	Documented In-House Method NEBBCRT/7 based on BS EN ISO 14644-1:1999:Part 1 and BS EN ISO 14644-2:2000:Part 2	B
	Cleanroom Cleanliness Classification Testing - Airborne Particulate Count		



1634
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

VEGA Environmental Consultants Ltd
Issue No: 013 Issue date: 03 September 2007

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ENVIRONMENTAL SAMPLES (cont'd) HEPA-ULPA filters	<u>Health and Hygiene</u> (cont'd) Filter Installation Leak Testing	NEBBCRT/6 based on BS PD 6609:2000 BS EN ISO 14644 - 3 and IES-RP-CC 006.2	A, B
END			